



Pollution Control Board:: Assam
Bamvnimaidam : Guwahati-21
Website: www.pcbassam.org

No. WB/LB-94/Pt-I/15-16/79

Dated Guwahati 19th August, 2016

SHORT QUOTATION NOTICE

Sealed quotations are hereby invited from reputed Manufacturer/Authorized Dealer/Supplier for the supply of Water Distillation Unit to the Pollution Control Board, Assam, Bamunimaidam, Guwahati-781021. The details Specification and the Terms & Conditions will be available on Board's Website: www.pcbassam.org/Tender.html from 20.08.2016.

The quotation will be received upto 14:00 hrs. of 31.08.2016 and will be opened on the same day at 16:00 Hrs 31.08.2016 in presence of the quotationers or their representatives.

Member Secretary (i/c)

02

Specification and the Terms & Conditions

Quotation Notice No. WB/LB-94/Pt-I/15-16/79 dtd. 19th August, 2016

Name of the Equipment	: Water Distillation Unit with DAPS & Water Softener (Mono/Double)
Material	: Quartz
Distilled water output capacity	: 2 to 3 Lts/Hr.
Conductivity ($\mu\text{S/cm}$)	: $<2 \times 10^{-6}$
Biological Activity	: Pyrogen Free
Power	: Single Phase, 230-240 Volts
Warranty	: 2 Years from the date of installation

Terms & Conditions:

1. Rate quoted should be inclusive of all taxes and F.O.R. destination at Board's Central Laboratory at Bamunimaidam, Guwahati-781021. Quoted rates should be both in figures and in words.
2. Delivery should be made within 4 to 6 weeks from the date of receipt of confirm supply order.
4. The bidder should be a manufacture/authorized dealer or supplier and must submit Authorization Certificate with the tender documents in case of dealer or supplier.
5. Operational, maintenance manual, warranty and certificates etc. are to be provided along with the instrument.
6. Safe transportation of the instrument up to Board's Central Laboratory is at risk of the supplier.
7. Prompt after sales services will have to be rendered as and when necessary.
8. Payment will be made only after receipt of the instrument in good condition and successful installation and commissioning of the instrument.
9. The Board at its discretion may extend the last date of submission of tender and opening of tenders. The final authority of acceptance of tender will rest with the Chairman, Pollution Control Board, Assam who does not bind himself to accept the lowest tender.

Member Secretary (i/c)

an