



Pollution Control Board, Assam

(Department of Environment & Forests :: Government of Assam)

অসম প্ৰদূষণ নিয়ন্ত্ৰণ পৰিষদ

(অসম চৰকাৰৰ বন আৰু পৰিৱেশ বিভাগ)

(An ISO 9001:2008 & BS OHSAS 18001:2007 Standard Certified Organisation)

No. WB/LAB-291/16-17/278

Dated Guwahati, the 29th July, 2017

Addendum-I to Board Tender Notice No.WB/LAB-291/16-17/233/832 dtd.29.07.2017.

In continuation of the Board tender notice No.WB/LAB-291/16-17/233/832 dtd.29.07.2017 for the supply Installation, Commissioning and O&M services of CAAQMS station at Guwahati following points are added to the bidding document.

1. As the proposed site for CAAQMS is on the 3rd floor of the PCBA Head Office building, the container for CAAQMS is not necessary and hence option of container in place of housing has been omitted from the bidding document (Volume-I). The civil construction of the room shall be done by the Board as per bidder diagram.
2. Qualifying requirement of the bidders (point No 2.5.6 page No.12) in the bidding document (Volume-I), trained O&M personal with following experience may also be consider.

Sl. No.	Responsibilities	No of Person	Minimum Qualification	Experience
1	Project Manager	1	Graduate Engineer/ M.Sc	10 years
2	Technician	1	Diploma in Electronics / C&I / Electrical	3 years

3. Board has received several quarries regarding single analyser for NO_x and NH₃ from bidders. Board is of the view that two separate analyser for the NH₃ & NO_x should be provided to minimise the data loss due to technical faults of the equipment if occurred. Thus the specification in the bidding document will remain the same.
4. Separate metrological sensors of wind speed / wind direction, Temperature and RH sensor are required.
5. Earnest money deposited (EMD) should be submitted in form of Demand Draft/ Banker Cheque only. BG instead of D.D shall not be accepted. The clause No. 4.4, page No.17, volume-I in the bidding document shall remain the same.

Yours faithfully,

Member Secretary (i/c)

Dated Guwahati, the 29th July, 2017

Memo No. WB/LAB-291/16-17/278-A

Copy to:

1. P.A. to the Chairman, PCBA for kind appraisal of the Hon'ble Chairman.
2. The Chief Env. Scientist, Central Laboratory, PCBA for favour of information.
3. The EITO of the Board to upload the Addendum in the Board Website.
4. Accounts Branch, PCBA for information & necessary action.

Member Secretary (i/c)

Head Office : Bamunimaidam, Guwahati - 781021, Assam : India.

Phone : 2652774 & 2550258 ; Fax : 0361-2550259 ; Gram : POLLUTIONCONTOL

E-mail : membersecretary@pcbassam.org; Website : www.pcbassam.org

Regional Offices at : Dibrugarh, Golaghat, Sibsagar, Tezpur, Guwahati, Bongaigaon, Nagaon & Silchar.